


Instructions for Radiation Implant Briefs Item 970-050 through 970-058


These briefs were designed for use with the Low Dose Rate (LDR) vaginal cylinder implant as well as the High Dose Rate (HDR) vaginal STUMP cylinder.


The Radiation Implant Brief is sized according to the patient's usual underwear size.


For LDR Vaginal Cylinder Implants




 Upon completion of the bowel cleanout, place the Radiation Implant Brief on the patient.







- The stick of the cylinder is inserted through the slit of the strap (B), and then the strap is secured to the top of the brief at the patient's waist at (A). This minimizes any mobilization of the cylinder.
- 5. For LOW DOSE RATE brachytherapy, the 3" outermost strap (C) (without a slit) should be brought up through the patient's legs. The desired degree of tension to keep the cylinder at the vaginal apex should then be applied to the stick of the cylinder, with the strap (C) then secured to the top of the pant over the underlying 2" strap.



6. For HDR Vaginal STUMP cylinders, place the brief on the patient. Through the 6" slit in the crotch of the brief, the foley catheter, as well as a rectal catheter may be inserted. Bring the innermost strap with the slit (B), from between the legs up over the crotch of the brief. The small circular stump of the cylinder fits in the slit of the strap. Secure strap (B) to the top of the brief, and then secure strap (C), without the slit, to the top of the brief over strap (B).

Patients having several **HDR** procedures may take the brief home and launder it in a regular washer/dryer. The same brief should not be used for different patients.

5218 Barthel Industrial Drive NE – Albertville, MN 55301 Phone: 763-497-2071 or 800-497-2071 Fax: 763-497-2295 www.rpdinc.com sales@rpdinc.com

G:\WORK\INFO\900\970\970-050-A Radiation Implant Brief Instructions.doc