



Expect Service

Radiation Products Design Inc

INSTRUCTIONS

RPD INFORMATION

Address	5218 Barthel Industrial Drive Albertville, MN 55301
Website	www.rpdinc.com
Email	sales@rpdinc.com
Phone	763-497-2071 or 800-497-2071
Fax	763-497-2295

RPD PRODUCT INFORMATION

Item Number	Description
260-301	Lung Compressor, Air Compression Belt w/Gauge, Small Compression area: 11.5 cm x 22.5 cm 4" wide Velcro Belt with tape CM/IN
260-302	Lung Compressor, Air Compression Belt w/Gauge, Large Compression area: 15 cm x 31 cm 6" wide Velcro Belt with tape CM/IN

DISCLAIMER

THESE PRODUCTS ARE NOT STERILE AND ARE TO BE USED BY AUTHORIZED PERSONNEL ONLY.

RADIATION PRODUCTS DESIGN INC assumes no liability for consequential damages of any kind for this material when used interchangeably with products of other manufacturers/suppliers or for any direct or indirect results and consequences of its use or misuse by the customer. Federal law (USA) restricts the sale of this device for use only by (or at the order of) a physician.

INTRODUCTION

The Lung Compressor, Air Compression Belt with Gauge is designed to put pressure just below the lungs to limit lung movement with respirations. The air bladder with a cotton cover is attached to a 60" (152.4 cm) long Velcro belt which is wrapped snugly around the patient just below the sternal notch. Once the belt and air bladder are in position on the patient, the air bladder is manually inflated with a hand pump. The included pressure gauge is used to record the amount of pressure used to achieve the desired limits of lung movement. This allows for reproducibility of the same amount of pressure for each treatment fraction.

CLEANING

This item can be cleaned by hand using a cloth dampened with a mild detergent dissolved in water. **DO NOT SOAK MATERIAL.** For stubborn spots or stains, use a soft brush and the solution. Allow material to air dry fully before reuse.

NOTE

Over inflation of the air bladder can cause seaming of the cotton cover to rip. To prevent air bladder cover from ripping, retighten belt to prevent over inflation of bladder.

End of Document