


TBI ROTATIONAL STAND HARNESS

Item# 495-6066



- A. HANGING CLIP
- B. VERTICAL STRAP
- C. CHEST STRAP
- D. LEG STRAPS (QTY. 2)

Directions:

Harness can also be worn reversed with the vertical strap and buckles in front when treating the back

Note: The folded face of the stitched should be facing away from the patient's skin to avoid chafing

1. Connect Hanging Clip (A) to TBI rotating hand rail steel handle
2. Connect Leg Straps (D) through the two loops on the bottom of the Vertical Strap (B), snug up, and drape Vertical Strap (B) over shoulder
3. Feed the Chest Strap (C) through the Vertical Strap (B) in the loop corresponding to the patient's underarm height and snug with the buckle behind the patient
4. Patient then stands on TBI platform
5. Connect the Vertical Strap (B) to the Hanging Clip (A) then tighten so the patient will remain in an upright position
6. Check that all belts are comfortable to patient, try to keep buckles out of radiation field

Cleaning Instructions:

The item can be cleaned by hand using a cloth dampened with a mild detergent dissolved in water. For stubborn spots, use a brush and the solution. Allow material to air dry before re-use.
DO NOT SOAK MATERIAL

